



# Sherwood Scientific Ltd.

1 The Paddocks, Cherry Hinton Road, Cambridge, UK. CB1 8DH

Tel: +44 (0) 1223 243444

Fax: +44 (0) 1223 243300



## DECLARATION OF CONFORMITY 08 February, 2010

We, Sherwood Scientific Ltd, UK, declare that the product(s) listed below:

|                                |                             |
|--------------------------------|-----------------------------|
| <i>Flame Photometer</i>        | <i>Model 360</i>            |
| <i>Flame Photometer</i>        | <i>Model 410 Classic</i>    |
| <i>Flame Photometer</i>        | <i>Model 410 Industrial</i> |
| <i>Flame Photometer</i>        | <i>Model 410 Clinical</i>   |
| <i>Flame Photometer</i>        | <i>Model 420 Industrial</i> |
| <i>Flame Photometer</i>        | <i>Model 420 Clinical</i>   |
| <i>Flame Photometer</i>        | <i>Model 425</i>            |
| <i>Continuous Flow Dilutor</i> | <i>Model 805</i>            |
| <i>Autosampler</i>             | <i>Model 860</i>            |

is/are in compliance with the essential requirements of the following EU Directives:

|                    |   |
|--------------------|---|
| <i>2006/95/EC</i>  | <i>Low Voltage Directive (LVD)</i>                        |
| <i>2004/108/EC</i> | <i>Electromagnetic Compatibility Directive (EMC)</i>      |
| <i>98/79/EC</i>    | <i>In-Vitro Diagnostic Medical Device Directive (IVD)</i> |

and has/have been designed and manufactured to conform to the essential requirements and specifications of the following harmonized standards:

|                            |   |
|----------------------------|---|
| <i>EN 61010-1:2001</i>     | <i>LVD – General Requirements</i>                             |
| <i>EN 61010-2-051:2003</i> | <i>LVD – Particular requirements for laboratory equipment</i> |
| <i>EN 61000-6-1:2007</i>   | <i>EMC - Generic standards - Immunity</i>                     |
| <i>EN 61000-6-3:2007</i>   | <i>EMC - Generic standards - Emission</i>                     |

.....  
Authorised Signatory  
Mr. N.J. Sugden - General Manager